

File : Q:\GCMS-08 Minimal\DATA\20230731\15- (JP).D
Operator : Organic QC Lab
Acquired : 31 Jul 23 04:27 pm using AcqMethod Hy_PAH_24cm_S150.M
Instrument : GCMS 8
Sample Name: AS-E0037
Misc Info : Fluoranthene; in meoh
Vial Number: 8



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

