

File :D:\MassHunter\GCMS\1\DATA\041122\43- SAK.D  
Operator : Organic QC Lab  
Acquired : 12 Apr 2022 08:26 using AcqMethod CHICK\_2019\_S100.M  
Instrument : GCMS 6  
Sample Name: AS-E0040  
Misc Info :  
Vial Number: 129

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25 $\mu$ m  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1 $\mu$ l inj.; GC/MS; Injector 270°C

