

File :Z:\QC Data\GCMS-07 Minimal\DATA\01092001\26-.D
Operator : Organic QC department
Acquired : 09 Jan 2020 20:18 using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name AS-E0058
Misc Info in MeOH
Vial Number: 88



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

