

File : L:\QC Data\GCMS-09 Minimal\DATA\08292022\06 - SAK.D  
Operator :  
Acquired : 29 Aug 2022 09:05 using AcqMethod Hy SV DB-5MS UI20M 0.18 40 split Time 12 min .M  
Instrument : GCMS 9  
Sample Name : AS-E0066  
Misc Info : Di-n-butyl phthalate in MeOH  
Vial Number : 27



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm  
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

