


File : Q:\GCMS-06 Minimal\DATA\111820\27- -DC.D
Operator : Organic QC Lab
Acquired : 18 Nov 2020 22:02 using AcqMethod CHICK_2017_13MIN_S100.M
Instrument : GCMS 6
Sample Name : AS-E0072
Misc Info : Benzo(b)fluoranthene in Acetone
Vial Number : 10

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 100:1, 1uL inj.; GC/MS; INJ 270c

