

File : L:\QC Data\GCMS-06 Minimal\DATA\032922\28- SAK.D  
Operator : Organic QC Lab  
Acquired : 29 Mar 2022 17:17 using AcqMethod CHICK\_2019.M  
Instrument : GCMS 6  
Sample Name: AS-E0129-2  
Misc Info : Arcolor in Isooctane  
Vial Number: 5

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

