

File :Q:\GCMS-08 Minimal\DATA\20220111\28- YX.D
Operator : Organic QC Lab
Acquired : 11 Jan 2022 22:53 using AcqMethod Hy_M502_S100.M
Instrument : GCMS 8
Sample Name: AS-E0252
Misc Info : p-Cresol in MeOH
Vial Number: 136



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25 μ m
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1 μ l inj.; GC/MS; Injector 270°C

