


File : Q:\GCMS-06 Minimal\DATA\020821\02- -DC.D  
Operator : Organic QC Lab  
Acquired : 08 Feb 2021 11:47 using AcqMethod CHICK\_2017\_13MIN\_S100.M  
Instrument : GCMS 6  
Sample Name: AS-E0261  
Misc Info : Dibenzofuran in MeOH  
Vial Number: 17

 **AccuStandard®**  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 100:1, 1uL inj.; GC/MS; INJ 270c

