


File :Q:\GCMS-07 Minimal\DATA\01032201\34- -KD.D
Operator : Organic QC department
Acquired : 04 Jan 2022 00:05 using AcqMethod KERLA-2017_Short_15Min_S100.M
Instrument : 5977B MSD
Sample Name: AS-E0369
Misc Info : Saccharin in AcCN
Vial Number: 19

 **AccuStandard**[®]
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

