


File : L:\QC Data\GCMS-06 Minimal\DATA\111819\10B-.D  
Operator : ZA / KM / YX  
Acquired : 18 Nov 2019 15:32 using AcqMethod CHICK\_2017\_VOC-SV\_S200.M  
Instrument : GCMS 6  
Sample Name: AS-E0480  
Misc Info : p-Dioxane  
Vial Number: 131

 **AccuStandard®**  
Leader in Analytical Reference Standards  
Column: DB-5MS, 30m, 0.25 ID, 0.25 um  
Oven Program: 80c, 17c/min to 340c, 8min  
GC Parameters: Cons. Split, 12psi constant flow  
Split 20:1, 1uL inj.; GC/MS; INJ 270c

