

File :D:\MassHunter\GCMS\1\DATA\092122\03- ALR.D  
Operator : Organic QC Lab  
Acquired : 21 Sep 2022 07:25 using AcqMethod CHICK\_2019\_13MIN.M  
Instrument : GCMS 6  
Sample Name: AS-E0797  
Misc Info : gamma-Chlordane in AcCN  
Vial Number: 36



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

