

File :D:\MassHunter\GCMS\1\data\11152022\03 - sak.D
Operator :
Acquired : 15 Nov 2022 08:45 using AcqMethod Hy VOC 24 AMU DB-5MS UI20M 0.18 50 split .M
Instrument : GCMS 9
Sample Name: AS-E1181
Misc Info :
Vial Number: 119



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

