

File :D:\MassHunter\GCMS\1\DATA\04112202\07- KD.D
 Operator : Organic QC department
 Acquired : 11 Apr 2022 11:47 using AcqMethod KERLA-2017.M
 Instrument : 5977B MSD
 Sample Name: ASM-184
 Misc Info :
 Vial Number: 64

Column: DB-5MS, 30m, 0.25 ID, 0.25 um
 Oven Program: 80c, 17c/min to 340c, 8min
 GC Parameters: Cons. Split, 12psi constant flow
 Split 20:1, 1uL inj.; GC/MS; INJ 270c

