


File :D:\MassHunter\GCMS\1\DATA\012221\06- -RS -DC.D
Operator : Organic QC Lab
Acquired : 22 Jan 2021 10:57 using AcqMethod CHICK_2019_13MIN.M
Instrument : GCMS 6
Sample Name: B-101S
Misc Info : Resample of B-101 @35ug/mL in Isooctane
Vial Number: 6

 **AccuStandard®**
Leader in Analytical Reference Standards
Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

