

File :Q:\GCMS-08 Minimal\DATA\20230106\18- [JP].D
Operator : Organic QC Lab
Acquired : 06 Jan 23 05:44 pm using AcqMethod Hy_BDE_24cm_S1.M
Instrument : GCMS 8
Sample Name: BDE-033S
Misc Info : 2'34-Tri-BDE; in iso
Vial Number: 19



Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

