

File : Q:\GCMS-08 Minimal\DATA\20230410\49- {JP=.D
Operator : Organic QC Lab
Acquired : 11 Apr 23 05:08 am using AcqMethod Hy_BDE_24cm_S1.M
Instrument : GCMS 8
Sample Name: BDE-066S
Misc Info : 23'44'-Tetra-BDE; in iso
Vial Number: 21



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj; GC/MS; Injector 270°C

