

File : L:\QC Data\GCMS-06 Minimal\DATA\082223\36 - SAK.D
Operator : Organic QC Lab
Acquired : 23 Aug 2023 05:36 using AcqMethod CHICK_2019_13MIN_2split_2uL.M
Instrument : GCMS 6
Sample Name : BDE-118S
Misc Info :
Vial Number : 28



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

