

File :Q:\GCMS-08 Minimal\DATA\20201207\58-YX.D
Operator : Organic QC Lab
Acquired : 08 Dec 2020 14:40 using AcqMethod Hy_BDE_S1.M
Instrument : GCMS 8
Sample Name: BDE-165S
Misc Info : 2,3,3',5,5',6-Hexabromodiphenyl ether
Vial Number: 2



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

