

Acq. Operator : SYSTEM  
 Acq. Instrument : LCMS-1  
 Injection Date : 7/13/2020 11:43:46 AM

Seq. Line : 15  
 Location : P1-B5  
 Inj : 1  
 Inj Volume : 1.000 µl

Method : C:\CHEM32\1\METHODS\Purity195\_315.M  
 Last changed : 5/8/2020 11:20:44 AM by SYSTEM  
 Method Info : Operator: Oraganic QC lab  
 Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 %  
 Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate  
 :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron  
 particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole  
 MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

