

File : L:\QC Data\GCMS-06 Minimal\DATA\061523\42 - sak.D
Operator : Organic QC Lab
Acquired : 15 Jun 2023 23:42 using AcqMethod CHICK_2017_13MIN_S100.M
Instrument : GCMS 6
Sample Name : BPA-C2-S
Misc Info :
Vial Number : 124



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 100:1, 1µl inj.; GC/MS; Injector temp 270°C

