

File :Q:\GCMS-08 Minimal\DATA\20210511\71- YX.D
Operator : Organic QC Lab
Acquired : 12 May 2021 15:46 using AcqMethod Hy_PAH_24cm_S50.M
Instrument : GCMS 8
Sample Name: BPA-E-S
Misc Info : Bisphenol in MeOH
Vial Number: 116



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

