

File : L:\QC Data\GCMS-09 Minimal\DATA\09262022\16 -SAK.D
Operator :
Acquired : 26 Sep 2022 15:21 using AcqMethod Hy SV DB-5MS UI20M 0.18 2 split FAST .M
Instrument : GCMS 9
Sample Name: C-009S
Misc Info :
Vial Number: 101



Column: DB-5MS UI, 20m x 0.18mm ID x 0.18µm
Oven Program: 45°C, 13°C/min to 290°C (hold 5min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector temp 270°C

