

File :D:\MassHunter\GCMS\1\data\20230413B\56- [JP].D
Operator : Organic QC Lab
Acquired : 14 Apr 23 01:13 pm using AcqMethod Hy_Pest_15cm_S10.M
Instrument : GCMS 8
Sample Name: C-030S-TP
Misc Info : 246-Tri-PCB; in iso
Vial Number: 41



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

