

File :D:\MassHunter\GCMS\1\data\20230405\46- \_JP).D  
Operator : Organic QC Lab  
Acquired : 06 Apr 23 02:50 pm using AcqMethod Hy\_Pest\_15cm\_S10.M  
Instrument : GCMS 8  
Sample Name: C-110S-TP  
Misc Info : 233'4'6-Penta-PCB; 100 ug/ml in iso  
Vial Number: 147



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

