

File :M:\GCMS-07 Minimal\DATA\05112101\32-ALR.D
Operator : Organic QC department
Acquired : 11 May 2021 17:42 using AcqMethod KERLA-2019_Short_15Min_2S_2uL.M
Instrument : 5977B MSD
Sample Name: C-119S
Misc Info : RT: 2,3',4,4',6-Pentachlorobiphenyl
Vial Number: 58



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

