

File :Q:\GCMS-06 Minimal\DATA\080521\44- -DC.D
Operator : Organic QC Lab
Acquired : 06 Aug 21 08:05 am using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: C-120S-TP
Misc Info : 2,3',4,5,5'-Pentachlorobiphenyl in Isooctane
Vial Number: 19



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

