

File :D:\MassHunter\GCMS\1\DATA\05112101\24-LR.D
Operator : Organic QC department
Acquired : 11 May 2021 15:03 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name:C-138S-TP
Misc Info : RT: 2,2',3,4,4',5'-Hexachlorobiphenyl in Isooctane
Vial Number: 44



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

