

File : M:\GCMS-07 Minimal\DATA\08042101\49- ALR.D  
Operator : Organic QC department  
Acquired : 05 Aug 2021 01:57 using AcqMethod KERLA-2019\_Short\_15Min\_2S\_2uL.M  
Instrument : 5977B MSD  
Sample Name : C-138S  
Misc Info : 2,2',3,4,4',5'-Hexachlorobiphenyl in Isooctane  
Vial Number: 64



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

