

File :M:\GCMS-06 Minimal\DATA\112321B\38- ALR.D  
Operator : Organic QC Lab  
Acquired : 24 Nov 2021 04:35 using AcqMethod CHICK\_2019\_13MIN\_2split\_2uL.M  
Instrument : GCMS 6  
Sample Name: C-153S  
Misc Info : 2,2',4,4',5,5'-Hexachlorobiphenylin Isooctane  
Vial Number: 15



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

