

File : L:\QC Data\GCMS-08 Minimal\DATA\20220520\16 - SAK.D
Operator : Organic QC Lab
Acquired : 20 May 2022 13:37 using AcqMethod Hy_Pest_15cm_S5.M
Instrument : GCMS 8
Sample Name : C-1555
Misc Info :
Vial Number : 17



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

