

File :Q:\GCMS-07 Minimal\DATA\01142101\38- 3 -DC.D
Operator : Organic QC department
Acquired : 15 Jan 21 12:51 am using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: C-180S-TP
Misc Info : C-180 in Isooctane
Vial Number: 16



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

