

File :Q:\GCMS-08 Minimal\DATA\20210820\50- (JP).D
Operator : Organic QC Lab
Acquired : 21 Aug 21 11:45 am using AcqMethod Hy_PCB_5cm_S5.M
Instrument : GCMS 8
Sample Name: C-182S-TP
Misc Info : 22'344'56'-Hepta-PCB in isoctane
Vial Number: 80



Column: DB-5, 30m, 0.25ID, 0.25um
Oven: 80c, 2m, 30c/m to 320, final 2min
GC para: Pulse split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj, GCMS, Inj 280c

