

File :Q:\GCMS-08 Minimal\DATA\20210713\25- YX.D  
Operator : Organic QC Lab  
Acquired : 13 Jul 2021 23:40 using AcqMethod Hy\_Pest\_15cm\_S2\_3ulinj.M  
Instrument : GCMS 8  
Sample Name: C-195S  
Misc Info : 2,2',3,3',4,4',5,6-Octachlorobiphenyl  
Vial Number: 133



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5m in, 10psi constant flow  
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

