

File :M:\GCMS-07 Minimal\DATA\07142101\48- ALR.D
Operator : Organic QC department
Acquired : 15 Jul 2021 01:16 using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: C-198S-TP
Misc Info : 2,2',3,3',4,5,5',6-Octachlorobiphenyl in Isooctane
Vial Number: 105



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

