

File :D:\MassHunter\GCMS\1\DATA\021422\17- ALR.D
Operator : Organic QC Lab
Acquired : 14 Feb 2022 15:22 using AcqMethod CHICK_2019_2S_2uL.M
Instrument : GCMS 6
Sample Name: C-200S-R1
Misc Info : 2,2',3,3',4,5,6,6'-Octachlorobiphenyl in Isooctane
Vial Number: 3



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

