

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-08 Minimal\DATA\20230328
... a\06- KM.D
Operator : Organic QC Lab
Instrument : GCMS 8
Acquired : 28 Mar 2023 18:28 using AcqMethod Hy_Pest_15cm_S10.M
Sample Name: C-221S-H
Misc Info : Aroclor 1221 in Hexane



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

