

File :D:\MassHunter\GCMS\1\data\20220126\32- ALR.D
Operator : Organic QC Lab
Acquired : 26 Jan 2022 19:58 using AcqMethod Hy_Pest_24cm_S1.M
Instrument : GCMS 8
Sample Name: C-242S-TP
Misc Info : Aroclor 1242 in Isooctane
Vial Number: 4



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

