

File :Q:\GCMS-08 Minimal\DATA\20191009\K15- D
Operator : KM / JP
Acquired : 10 Oct 2019 10:03 using AcqMethod KERLA-2017_S100.M
Instrument : GCMS 8
Sample Name: C-260-ST-1 dil. 10:1 in Isooctane
Misc Info : Aroclor 1260 in Transformer Oil dil. 10:1 in Isooctane
Vial Number: 37



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

