

File : L:\QC Data\GCMS-08 Minimal\DATA\20220422\04 - SAK.D
Operator : Organic QC Lab
Acquired : 22 Apr 2022 11:06 using AcqMethod Hy_Pest_15cm_S50.M
Instrument : GCMS 8
Sample Name: C-268S-H-10X
Misc Info :
Vial Number: 44

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

