File :D:\MassHunter\GCMS\1\data\05271901\10- D
Operator : KD
Acquired : 28 May 19 10:36 am using AcqMethod KERLA-2Split-2uL.M
Instrument : 5977B MSD
Sample Name: C-CS-06
Misc Info : PCB Congener Mix
Vial Number: 22

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.