

File :Q:\GCMS-08\DATA\20190306\BSB\05- .D
Operator : JP/KM
Acquired : 06 Mar 19 05:45 pm using AcqMethod KERLA-2Split-2uL.M
Instrument : GCMS 8
Sample Name: C-WHO-01
Misc Info : PCB Mix in Isooctane
Vial Number: 15



AccuStandard®

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

