



=====

Acq. Operator : SYSTEM	Seq. Line : 8
Acq. Instrument : LCMS-1	Location : P1-A9
Injection Date : 9/5/2019 5:43:48 PM	Inj : 1
	Inj Volume : 1.000 µl

Acq. Method : D:\DATA\09051901\09051901 2019-09-05 15-32-34\Purity195_315.M

Last changed : 9/5/2019 3:32:34 PM by SYSTEM

Analysis Method : C:\Chem32\1\Methods\Purity195_315.M

Last changed : 4/10/2019 10:56:50 AM by SYSTEM

Method Info

: Operator: GA

Gradient Method Starting @ 100 % Water/0 % Acetonitrile to 0 % Water/100 % Acetonitrile in 13 min, then 3 min at 100/0, post-run time 2 min. Flow rate :0.3 ml/min, column Zorbax Eclipse Plus C-18, 2.1x50 mm, 1.8 micron particule size. Agilent 1290 Infinity II LC system, Agilent 6120 Quadropole MS Detector. No UV Reference.

Additional Info : Peak(s) manually integrated

Current Chromatogram(s)

