

File : Q:\GCMS-07\DATA\01181901\13-
Operator : K.D
Acquired : 18 Jan 2019 12:55 using AcqMethod KERLA-2017_S100.M
Instrument : 5977B MSD
Sample Name: CLP-011B
Misc Info : in CH2Cl2
Vial Number: 17

Column: DB-5, 30m, 0.25 ID, 0.25um
Oven Program: 100c, 1m, 30c/min to 320c, final 2min.
GC Parameters: Pulsed split 40psi 0.5m, 10psi constant flow
Split 20:1, 1ul inj.; GC/MS; INJ 250c.

