

File : Q:\GCMS-07 Minimal\DATA\05282001\44- DC.D
Operator : Organic QC department
Acquired : 29 May 20 02:58 am using AcqMethod KERLA-2017.M
Instrument : 5977B MSD
Sample Name: CLP-014-5X-25ML
Misc Info : Matric Std. Spiking Soln. in Methanol
Vial Number: 85



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 35c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

