

File : Q:\GCMS-04\DATA\122118\09-  
 Operator : RFXM  
 Acquired : 21 Dec 2018 19:43 using AcqMethod VOC-100SPLIT DECKER.M  
 Instrument : GCMS4  
 Sample Name: CLP-022  
 Misc Info : 1  
 Vial Number: 76



Column: DB-624 UI, 30m x 0.25mm ID x 1.4µm  
 Oven Program: 35°C (hold 5min), 11°C/min to 60°C, 22°C/min to 230°C (hold 4min)  
 GC Parameters: Split 100:1, 1µl inj.; GC/MS; Injector temp 240°C

