

File :Q:\GCMS-08\DATA\20190104\07-
Operator : JJP / KM
Acquired : 04 Jan 2019 14:08 using AcqMethod KERLA-2017_S50.M
Instrument : GCMS 8
Sample Name: CLP-027-R2 dil. 10:1 in MeCl2
Misc Info : GPC Cal. Standard in MeCl2
Vial Number: 16



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 20:1, 1µl inj.; GC/MS; Injector 270°C

