

File :Q:\GCMS-08 Minimal\DATA\20230110\07- (JP).D
Operator : Organic QC Lab
Acquired : 10 Jan 23 11:00 am using AcqMethod Hy_PAH_24cm_S20.M
Instrument : GCMS 8
Sample Name: CLP-032-R
Misc Info in acetone
Vial Number: 14



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

