

File :L:\QC Data\GCMS-06 Minimal\DATA\113022\14 SAK.D
Operator : Organic QC Lab
Acquired : 30 Nov 2022 15:25 using AcqMethod CHICK_2019.M
Instrument : GCMS 6
Sample Name: CLP-034
Misc Info :
Vial Number: 107



Column: DB-5MS UI, 30m x 0.25mm ID x 0.25µm
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow
Split 2:1, 2µl inj.; GC/MS; Injector 270°C

