

File : \\SVR-FS-01\Organic QC\QC Data\GCMS-07 Minimal\DATA\02102302  
... \19- KM.D  
Operator : Organic QC department  
Instrument : 5977B MSD  
Acquired : 10 Feb 2023 15:36 using AcqMethod KERLA-2017\_Short\_15Min.M  
Sample Name: CLP-BNS-3-2X  
Misc Info : 4-Terphenyl-d14 in DCM



**AccuStandard<sup>®</sup>**

Column: DB-5MS UI, 30m x 0.25mm ID x 0.25 $\mu$ m  
Oven Program: 80°C, 17°C/min to 340°C (hold 8min)  
GC Parameters: Pulsed split 40psi 0.5min, 10psi constant flow  
Split 20:1, 1 $\mu$ l inj.; GC/MS; Injector 270°C

