

File : Q:\GCMS-07 Minimal\DATA\02232101\38- -DC.D
Operator : Organic QC department
Acquired : 23 Feb 21 08:43 pm using AcqMethod KERLA-2017_Short_15Min.M
Instrument : 5977B MSD
Sample Name: CLP-FC
Misc Info : 2,4,5-Trichlorophenol in Acetone
Vial Number: 138



Column: DB-5MS, 30m, 0.25 ID, 0.25 um
Oven Program: 80c, 17c/min to 340c, 8min
GC Parameters: Cons. Split, 12psi constant flow
Split 20:1, 1uL inj.; GC/MS; INJ 270c

